

Certificate of Analysis

Sample Information

ARL ID: 745193
Date Received: 5/10/2024
Date Tested: 5/15/2024
Description: Happy Sac
Lot#: 240426

Results

Analysis	Method	†MDL / LOQ	Specification	Results	UOM	Lab ID
<u>Complete Micro Profile Pseudomonas</u>	USP <2021>, USP <2022>, AOAC 991.14, USP <62>					24
Total Plate Count	USP <2021>	10	Record Only	None Detected	cfu's/g	24
Coliforms	AOAC 991.14	10	Record Only	None Detected	cfu's/g	24
E. coli	USP <2022>	Absent	Record Only	Absent	cfu's/10g	24
Staphylococcus aureus	USP <2022>	Absent	Record Only	Absent	cfu's/10g	24
Salmonella	USP <2022>	Absent	Record Only	Absent	cfu's/10g	24
Pseudomonas aeruginosa	USP <62>	Absent	Record Only	Absent	cfu's/g	24
Yeast	USP <2021>	10	Record Only	None Detected	cfu's/g	24
Mold	USP <2021>	10	Record Only	None Detected	cfu's/g	24

†Method Detection Limit (MDL):

In microbiological testing, this is the minimum level of growth that can be detected with confidence. If a result is reported as "None Detected", it means any visible growth was below this limit.

†Limit of Quantitation (LOQ):

In analytical chemistry testing, this is the minimum level of the desired analyte that can be quantified with confidence. If a result is reported as less than LOQ, it means any detected amount was too small to report an exact number.

Under accreditation number 77504, ARL is an ISO/IEC 17025:2017 Accredited Laboratory. Uncertainty data for ISO-scoped methods is available upon request. Certificate and scope are also available upon request.

